RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

TECHNICAL SERVICE BULLETIN # 9

Product:	Fabius GS
Date:	March 22, 2004
Re:	Software Release - 1.39.3
Reference Doc:	Product Bulletin #163, SP00312
Reason/Rational:	Enhancement – See Product Bulletin #163 for details
Solution/Procedure:	Update 1.39.1 (field) Software kits to 1.39.3. Perform Software 1.39.3 upgrade on all Fabius GS units containing 1.3n. Complete the Return Form (attached) and return to DMI/DS within 7 days of account completion.
Devices affected:	All units containing 1.3n **only.**
Timing/When:	At customer request, next service or Periodic Manufacturer's Service
Additional Info:	Software 1.39.3 supersedes1.39.1 upon release. Therefore, all 1.39.1 kits located in the field shall be updated to 1.39.3 by ordering P/N 4117937-010 - Software 1.39.3 CD. Or, future machine updates from 1.3n to 1.39.3 shall be completed using Software Kit – 4118733 Note Do not install software 1.39.3 on units containing 2.n. See TSB #10 for details.
Cost:	Software upgrades to 1.39.3 are available to Fabius GS customers at no charge.
Ordering Info/Parts:	Contact your DMI Customer Service Representative to order the software.
Distribution:	DrägerService N.A. & all Authorized Service Organizations.

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.

RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Product: Fabius GS
Re: TSB #9

Return Form - Software 1.39.3 Upgrade

Distributor:	Customer:
Country:	Address:
	City:

Machine S/N	Machine S/N	Machine S/N
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Fax To:
DrägerService
Attn: Al Myers
(215) 721-5784

Date of Update _____
Signature (TSR) _____
Name (Print) _____

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.